

γ -GlutamylTranspeptidase (γ -GT) Activity Assay Kit

Note: Take two or three different samples for prediction before test.

Operation Equipment: Spectrophotometer

Catalog Number: BC1220

Size: 50T/48S

Components

Extract solution: 50 mL \times 1. Storage at 4°C.

Reagent I: Powder \times 1. Storage at 4°C.

Reagent II: 12.5 mL \times 1. Storage at 4°C.

Reagent III: 44.5 mL \times 1. Storage at 4°C.

Working solution (prepare in Reagent I bottle): prepare when the solution will be used, pour the Reagent II into Reagent I bottle, fully dissolved (incubate in 40°C water bath to promote the dissolution if the room temperature is too low). Then pour Reagent III into Reagent I bottle, mix well and store at room temperature.

Product Description

γ -glutamyltranspeptidase (γ -GT) is a key enzyme in γ -glutanyl cycle, which catalyzes the degradation of GSH. γ -GT catalyzes the transfer of γ -glutamyl groups from GSH or other γ -glutamyl compounds to receptors. It can also catalyze the hydrolysis of GSH and other γ -glutamyl compounds to produce glutamate, which plays an important role in the metabolism of extracellular glutathione.

γ -GT catalyzes the transfer of γ -glutamyl in glutamyl p-nitroaniline to N-glycylglycine to form p-nitroaniline with characteristic light absorption at 405 nm. γ -GT enzyme activity is calculated by measuring the increase rate of light absorption at 405 nm.

Reagents and Equipment Required but Not Provided

Spectrophotometer, centrifuge, water-bath, adjustable pipette, 1 mL glass cuvette, ice, mortar/homogenizer and distilled water.

Procedure

I. Extraction of crude enzyme solution:

1. Bacteria or cultured cells:

Collect bacteria or cells into centrifuge tube, discard the supernatant after centrifugation. According to the number of bacteria or cells (10^4): the Extract solution volume (mL) is 500~1000:1 (it is recommended that add 1 mL of the Extract solution to 5 million bacteria or cells), break the bacteria or cells by ultrasound (ice bath, 20% power or 200W, ultrasound 3s, interval of 10s, repeat for 30 times). Centrifuge at 10000 rpm for 10 minutes at 4°C, take the supernatant and place it on ice for test.

2. Tissue:

Weigh about 0.1 g of samples, add 1.0 mL of extract solution, full grinding. Centrifuge at 10000rpm for 15 minutes at 4°C, take the supernatant and place it on ice for test.

3. Serum (plasma):

Direct detection.

II. Test Steps:

1) Preheat the Spectrophotometer for more than 30 minutes, adjust the wavelength to 405nm and set the zero with distilled water.

2) Place working solution at 25°C (general species) or 37°C (mammals) water bath, preheating for more than 30 minutes (Ensure that there is no precipitation).

3) Sample test:

Reagent(μL)	Blank Tube (A _B)	Test tube (A _T)
Distilled water	100	-
Supernatant/serum	-	100
Working solution	900	900

After mixing thoroughly, detect the absorbance value at 405nm at 10s(A₁) and 130s(A₂). Calculation:

$\Delta A = A_2 - A_1$. Calculate $\Delta A_T = \Delta A - \Delta A_B$.

III. Calculation of γ -GT activity

1. Calculate by sample protein concentration

Active unit (U) definition: One unit of enzyme activity is defined as the amount of enzyme catalyzes the produce of 1 μmol of P-nitroaniline per minute at 25°C or 37°C every milligram of protein.

γ -GT(U/mg prot) = $\Delta A_T \div (\epsilon \times d) \times 10^6 \times V_{TV} \div (C_{pr} \times V_S) \div T = 0.506 \times \Delta A_T \div C_{pr}$.

2. Calculate by sample fresh weight

Active unit (U) definition: One unit of enzyme activity is defined as the amount of enzyme catalyzes the produce of 1 μmol of P-nitroaniline per minute at 25°C or 37°C every gram of tissue.

γ -GT(U/g fresh weight) = $\Delta A_T \div (\epsilon \times d) \times 10^6 \times V_{TV} \div (W \div V_E \times V_S) \div T = 0.506 \times \Delta A_T \div W$.

3. Calculate by serum (plasma)

Active unit (U) definition: One unit of enzyme activity is defined as the amount of enzyme catalyzes the produce of 1 μmol of P-nitroaniline per minute at 25°C or 37°C every per liter of serum.

γ -GT(U/L serum (plasma)) = $\Delta A_T \div (\epsilon \times d) \times 10^6 \times V_{se(pla)} \div T = 0.506 \times \Delta A_T$.

4. Calculated by bacteria or cultured cells

Active unit (U) definition: One unit of enzyme activity is defined as the amount of enzyme catalyzes the produce of 1 μmol of P-nitroaniline per minute at 25°C or 37°C every ten thousand bacteria or cells.

γ -GT(U/10⁴ cell) = $\Delta A_T \div (\epsilon \times d) \times 10^6 \div (500 \times V_S \div V_E) \div T = 0.001 \times \Delta A_T$.

V_S: Add sample volume, 0.1mL;

V_E: Add extraction liquid volume: 1mL;

T: Reaction time, 2 minutes;

C_{pr}: Sample protein concentration, mg/mL;

W: Sample weight, g;

5 million:5 million cells;

ϵ : The extinction coefficient of P-nitroaniline is 9870 L/mol/cm;

d: Light path of cuvette, 1cm;

V_{TV} : Total volume of reaction system, 0.001L;

10^6 : Unit conversion coefficient, 1mol= $10^6\mu\text{mol}$;

$V_{se(pla)}$: Volume of serum (plasma), 0.1mL.

Note:

When measure the activity of γ -GT in cultured cells, the extraction process of γ -GT in cells could by grinding or ultrasonic treatment after adding reagent. Cells can not treat with cell lysis buffer (prevent the deactivation of enzymes due to protein degeneration).

Experimental instances:

1. Take 0.1g of kidney, add 1mL of extract solution, homogenate and grind. Centrifuge at 10000rpm for 10 minutes at 4°C, take the supernatant, dilute it by 20 times, and test according to the measured steps. Calculate $\Delta A_T = A_{T2} - A_{T1} = 1.412 - 0.68 = 0.732$, $\Delta A_B = A_{B2} - A_{B1} = 0.597 - 0.578 = 0.019$, calculate the enzyme activity according to sample weight:

$$\gamma\text{-GT (U/g weight)} = 0.506 \times \Delta A \div W \times 20 \text{ (dilution ratio)} = 72.16 \text{ U/g weight.}$$

Related products:

BC1170/ BC1175 Reduced Glutathione (GSH) Assay Kit

BC1180/BC1185 Oxidized Glutathione (GSSG) Assay Kit