

L-Lactate Dehydrogenase (L-LDH) Activity Assay Kit

Note: Take two or three different samples for prediction before test.

Operation Equipment: Spectrophotometer/microplate reader

Catalog Number: BC0685

Size:100T/48S

Product Composition: Before use, please carefully check whether the volume of the reagent is consistent with the volume in the bottle. If you have any questions, please contact Solarbio staff in time.

Re	agent Name	Size	Preservation Condition	
Extract solution		Liquid 60 mL×1	2-8°C	
ic	Reagent I	Liquid 7 mL×1	2-8°C	
Reagent II		Powder ×1	-20°C	
JUFE I	Reagent III	Liquid 7 mL×1	2-8°C	
I	Reagent IV	Liquid 25 mL×1	2-8°C	
	Standard	Liquid 1 mL×1	⊚ 2-8°C	

Solution Preparation:

1. Reagent II: powder $\times 1$ bottle, add 1.3 mL of distilled water before use. It can be divided into tubule after matching, the unused reagent can be stored at -20°C for 2 weeks, avoid repeated freezing and thawing.

2. Standard: liquid ×1 bottle, 20 µmol/mL Sodium pyruvate.

Product Description:

L-Lactate dehydrogenase (L-LDH or LD) is the terminal enzyme of the glycolysis pathway which is found in nearly all living cells (animals, plants, and prokaryotes). L-LDH catalyzes the conversion of lactate to pyruvic acid and back, as it converts NAD⁺ to NADH and back.

NAD⁺ and lactic acid are oxidized to pyruvic acid by the catalysis of L-LDH. Pyruvate further reacted with 2,4-dinitrophenylhydrazide to form pyruvate dinitrobenzone, which show brown red color in alkaline solution and the color depth is proportional to the concentration of pyruvate.

Note: Before the experiment, it is recommended to select 2-3 sample with large expected differences for pre-experiment. If the absorption value of the sample is not within the measurement range, it is recommended to dilute or increase the sample size for detection

Reagents and Equipment Required but Not Provided:

Spectrophotometer/Microplate reader, thermostat water bath, desk centrifuge, adjustable pipette, micro glass cuvette/96 well flat-bottom plate, mortar/homogenizer/cell ultrasonic crusher, ice, distilled water.

Operation procedure:

I. Sample Preparation.(The sample size to be tested can be adjusted appropriately, and the specific

BC0685 -- Page 1 / 4

Tel: 86-010-50973105 https://www.solarbio.net

proportion can be referred to the literature.)

1. Bacteria, cells or tissue sample:

Bacteria or cells:

Collecting bacteria or cells into the centrifuge tube. The liquid in the upper layer is discarded after centrifugation. The ratio of bacteria/cell amount (10⁴): Extract solution volume (mL) is 500~1000:1(it is suggested to take about 5 million bacteria/cell and add 1 mL of Extract solution). Bacteria and cell is split

by ultrasonic (placed on ice, 200W, work time 3s, interval 10s, repeat for 30 times). Centrifuge at 8000 rpm 4°C for 10 minutes, take the supernatant and put it on ice for test.

Tissue:

Ice-bath homogenate is conducted according to the ratio of tissue mass (g): Extract solution volume (mL) = 1: $5\sim10$ (it is suggested to take about 0.1 g of tissue and add 1 mL of Extract solution). Ice bath homogenization. Centrifuge at 8000 rpm and 4°C for 10 minutes, take the supernatant and put it on ice for test.

2. Serum (plasma) sample:

Detect sample directly.

Note: The sample homogenate supernatant of the kit is also available for BC0090/BC0095 (Peroxidase), BC0020/BC0025 (Malondialdehyde), BC0200/BC0205 (Catalase), BC0170/BC0175 (Superoxide dismutase), and BC5160/BC5165 (superoxide dismutase).

II. Determination procedure.

- 1. Preheat the Spectrophotometer/Microplate reader 30 minutes, adjust wavelength to 450 nm, set zero with distilled water.
- 2. 20 μmol/mL standard solution is respectively diluted to 2, 1, 0.5, 0.25, 0.125 and 0 mmol/mL with distilled water and use 2, 1, 0.5, 0.25, 0.125, 0 μmol/mL as standard curve.

Reagent name (µL)	Test tube(At)	Control tube(Ac)	Standard tube(As)
Sample	10	10	SUFES
Standard Solution	- joio	-	10
Reagent I	50	50	50
Reagent II	10	- 50	©
Distilled water	S -	10	٤ 10
Mixed thoroughly, in	cubate at 37°C(mamma	l) or 25°C(other species)	water bath for 15 minutes.
Reagent III	50	50	50
Mixed thoroughly, ine	cubate at 37°C(mamma	l) or 25°C(other species)	water bath for 15 minutes.
Reagent IV	150	150	150

3. Sample Test

Mixed thoroughly, place at room temperature for 3 minutes. Take 200 μ L of reaction solution in micro glass cuvette/96 well flat-bottom plate, measured the absorbance at 450 nm, $\Delta A=A_T-A_C$. Each test tube should set a control tube.

BC0685 -- Page 2 / 4



III. L-LDH Calculations

1. Set the standard curve, x-axis as the standard concentration, μ mol/mL, y-axis as the 450 nm absorption. Put $\Delta A(y)$ into standard curve, calculate x (μ mol/mL)

2. Sample Sodium pyruvate content

Put $\Delta A(y)$ into standard curve, calculate x (µmol/mL)

3. Serum (plasma) sample L-LDH activity

Unit definition: One unit of enzyme activity is defined as the amount of enzyme catalyzes the producing of 1 nmol of pyruvic acid per minute every milliliter of serum.

L-LDH(U/mL)= $x \times Vs \div Vs \div T \times 10^3$ =66.7×x

4. Tissue, bacteria or cultured cells L-LDH activity

A. Calculate by sample protein concentration

Unit definition: One unit of enzyme activity is defined as the amount of enzyme catalyzes the producing of 1 nmol of pyruvic acid per minute every milligram of protein.

L-LDH(U/mg prot)= $x \times Vs \div (Cpr \times Vs) \div T \times 10^3 = 66.7 \times x \div Cpr$

B. Calculate by sample mass

Unit definition: One unit of enzyme activity is defined as the amount of enzyme catalyzes the producing of 1 nmol of pyruvic acid per minute every gram of tissue.

L-LDH(U/g mass)= $x \times Vs \div (W \div Vsv \times Vs) \div T \times 10^3 = 66.7 \times x \div W$

C. Calculate by the number of bacteria or cells

Unit definition: One unit of enzyme activity is defined as the amount of enzyme catalyzes the producing of 1 nmol of pyruvic acid per minute every 1 0000 cells.

L-LDH (U/10⁴ cell) = $x \times Vs \div (N \div Vsv \times Vs) \div T \times 10^3 = 66.7 \times x \div N$

Vs: Supernate volume (mL), 10μ L = 0.01 mL;

Vsv: Extract solution volume, 1 mL;

T: Reaction time, 15 minutes;

Cpr: Sample protein concentration, mg/mL;

W: Sample weight, g

N: Total number of bacteria or cells, million;

 10^3 : 1 µmol/mL= 10^3 nmol/mL.

Note: When ΔA is greater than 1.3 or less than 0.01, it is recommended to dilute the sample with distilled water or increase the sample size for the experiment, and pay attention to the simultaneous modification of the calculation formula.

Experimental example:

1. Take 0.103g of seduma leaves, add 1mL of extract solution, make ice bath homogenate, 8000g, centrifuge at 4°C for 10min, take superclear ice to be measured. Then, according to the

BC0685 -- Page 3 / 4



measurement procedure, the 96-well plate was used to calculate $\Delta A=A$ measuring tube -A control =0.275-0.199=0.076,

and the standard curve y=0.5218x+0.0063, $R^2=0.9983$, x=0.134 µmol/mL, and the lactate dehydrogenase activity was calculated to obtain:

L-LDH (U/g mass) =66.67 ×x÷W =86.74 U/g

2. Take 0.109 g of rabbit liver, add 1mL of extract liquid, homogenize in ice bath, 8000g, centrifuge at 4°C for 10min, take supernatant and dilute 80 times with distilled water, then place on ice to be measured. Then, according to the measurement procedure, the 96-well plate was used to calculate $\Delta A=A$ measuring tube -A control = 0.7955-0.132 =0.663, and the standard curve y=0.5218x+0.0063, R²=0.9983, x=1.259 µmol/mL, and the lactate dehydrogenase activity was calculated to obtain:

L-LDH (U/g mass) =66.67 \times x \div W \times dilution ratio =61605.53 U/g

3. After taking 10µL equine serum, follow the measurement procedure, using A 96-well plate to calculate ΔA =A measuring tube -A control =0.415-0.161=0.254, and adding the standard curve y=0.5218x+0.0063, R²=0.9983, x=0.475 µmol/mL, and calculate the lactate dehydrogenase activity to obtain:

L-LDH (U/mL) =66.67 x x =31.67 U/mL

Recent Product Citations:

[1] Yang W, Liu LB, Liu FL, Wu YH, Zhen ZD, Fan DY, Sheng ZY, Song ZR, Chang JT, Zheng YT, An J, Wang PG. Single-cell RNA sequencing reveals the fragility of male spermatogenic cells to Zika virus-induced complement activation. Nat Commun. 2023 Apr 29;14(1):2476. doi: 10.1038/s41467-023-38223-z. PMID: 37120617; PMCID: PMC10148584.

[2] Xiao W, Chen W, Wang Y, Zhang C, Zhang X, Zhang S, Wu W. Recombinant DTβ4-inspired porous 3D vascular graft enhanced antithrombogenicity and recruited circulating CD93+/CD34+ cells for endothelialization. Sci Adv. 2022 Jul 15;8(28):eabn1958. doi: 10.1126/sciadv.abn1958. Epub 2022 Jul 13. PMID: 35857526; PMCID: PMC9278867.

[3] Yu Y, Tian D, Yu Y, Lu L, Shi W, Liu G. Microplastics aggravate the bioaccumulation and corresponding food safety risk of antibiotics in edible bivalves by constraining detoxification-related processes. Sci Total Environ. 2024 Jan 15;908:168436. doi: 10.1016/j.scitotenv.2023.168436. Epub 2023 Nov 8. PMID: 37949129.

References:

[1] Huang P H, Fu L C, Huang C S, et al. The uptake of oligogalacturonide and its effect on growth inhibition, lactate dehydrogenase activity and galactin-3 release of human cancer cells[J]. Food chemistry, 2012, 132(4): 1987-1995.

Related Products:

BC0740/BC0745	Hexokinase(HK) Activity Assay Kit
BC0540/BC0545	Pyruvate Kinase(PK) Activity Assay Kit

BC0685 -- Page 4 / 4

