

Glucose Content Assay Kit (O-toluidine Colorimetry)

Note: Take two or three different samples for prediction before test.

Operation Equipment: Spectrophotometer

Catalog Number: BC5870

Size:50T/48S

Components:

Extract Solution: 60 mL×1. stored at 2-8°C.

Reagent I: 6 mL×1, stored at 2-8°C. Take 9mL reagent II, 45mL acetic acid add reagent I mixed, the final working solution is a homogeneous clarified liquid. The dissolved reagent can be stored at -20°C for 8 weeks. (Note: If Reagent I and Reagent II are mixed first, the solution is a liquid mixture of inorganic and aqueous phases, and the solution will become homogeneous and clarified after adding glacial acetic acid.)

Reagent II:10mL×1, stored at 2-8°C.

Standard: Powder×1. Store at 2-8°C. Add 1 mL of distilled water to fully dissolve the 10 mg/mL glucose standard solution before use; store the inexhaustible reagent at 2-8°C for 4 weeks. Before use, take 30 μ L of 10 mg/mL glucose standard solution and add 930 μ L of distilled water to make 0.3125 mg/mL glucose standard solution.

Product Description

Glucose is not only the main substrate for cellular energy metabolism, but also its metabolic intermediates are important substrates for biosynthesis. In the case of mammals, glucose is not only the only energy source for the nervous system of the brain, muscles, adipose tissue, etc, but also is closely related to the synthesis of reducing coenzymes, lactose and milk fat.

When glucose and o-toluidine are heated in a strong acid solution, the aldehyde group of glucose condenses with o-toluidine to form glucosylamine, which dehydrates to form Schiff base, which is blue-green in color with an absorption peak at 630nm. **The kit is suitable for the determination of animal tissue, serum (plasma) and cell (bacterial) samples.**

Reagents and Equipment Required but Not Provided.

Visible spectrophotometer, cryogenic centrifuge, analytical balance, 1mL glass cuvette, mortar/homogenizer/cell ultrasonic breaker, adjustable pipette, acetic acid (AR > 98%), 5% trichloroacetic acid (determine if you need to bring your own depending on precautions), isopropyl alcohol (determine if you need to bring your own depending on precautions), ice and distilled water.

Procedure:

I. Sample Extraction:

1. Tissue treatment:

The tissue mass (g): volume of distilled water (mL) is 1:5-10 (it is recommended to weigh about 0.1g of tissue and add 1mL of distilled water), grind into a homogenate, boil in a boiling water bath for 10 min

(cover tightly to prevent water loss), cool to room temperature, centrifuge at 8000g for 10 min at 25°C, and remove the supernatant.

2. Bacteria or cell treatment:

Collect the bacteria or cells into the centrifuge tube, discard the supernatant after centrifugation; According to the bacteria or cells (10^4) : distilled water volume (mL) is according the ratio of 500~1000: 1 (Recommend 1 mL of distilled water is added to 5 million bacteria or cells), ultrasonic broke bacteria or cells (ice bath, power of 200W, ultrasound for 3s, interval of 10s, repeat 30 times), set in a boiling water bath boil for 10 minutes (tightly closed to prevent moisture loss), after cooling, 8000 g, 25°C centrifuge for 10 min, take supernatant on standby.

3. Liquid samples such as serum (plasma):

Direct measurement. If the liquid is turbid, centrifuge the supernatant.

II. Determination procedure:

1. Preheat the spectrophotometer for 30min, adjust the wavelength to 630nm and adjust zero with acetic acid.

2. Add the following reagents successively into the 1.5ml centrifuge tube:

Reagent (μL)	Test Tube (A _T)	Standard Tube (A _S)	Blank Tube (A _B)
Sample	100	-	-
Standard	-	100	-
ddH ₂ O	-	-	100
Reagent I	1000	1000	1000

Vortex mixing, boiling at 100 °C for 15min (cover tightly to prevent water loss), cooling to room temperature, take 1000μL to 1mL glass cuvette to determine the absorbance value at 630nm, were recorded as A_t, A_s, A_b. Calculate $\Delta A_t = A_t - A_b$ and $\Delta A_s = A_s - A_b$. The standard and blank tubes should be measured only 1-2 times.

III. Calculation:

1. Calculate by the protein concentration:

$$\begin{aligned} \text{Glucose content (mg/mg prot)} &= C_s \times \Delta A_t \div \Delta A_s \times V_s \div (C_{pr} \times V_s) \\ &= 0.3125 \times \Delta A_t \div \Delta A_s \div C_{pr} \end{aligned}$$

2. Calculate by Sample fresh weight:

$$\begin{aligned} \text{Glucose content (mg/g fresh weight)} &= C_s \times \Delta A_t \div \Delta A_s \times V_E \div W \\ &= 0.3125 \times \Delta A_t \div \Delta A_s \div W \end{aligned}$$

3. Calculate by volume of liquid

$$\begin{aligned} \text{Glucose content (mg/mL)} &= C_s \times \Delta A_t \div \Delta A_s \times V_s \div V_s \\ &= 0.3125 \times \Delta A_t \div \Delta A_s \end{aligned}$$

4. Calculate by the number of bacteria or cells

$$\begin{aligned} \text{Glucose content (mg/10}^4 \text{ cell)} &= C_s \times \Delta A_t \div \Delta A_s \times V_E \div N \\ &= 0.3125 \times \Delta A_t \div \Delta A_s \div N \end{aligned}$$

C_s: glucose solution concentration, 0.3125mg/mL;

C_{pr}: sample protein concentration, mg/mL;

V_S : the sample volume added, 100 μL =0.1 mL;

V_E : Volume of pre-treatment extract, 1 mL;

W: sample fresh weight, g;

N: number of bacteria or cells.

Note:

- 1、 Reagent I is corrosive, please wear gloves and mask and operate carefully.
- 2、 If the final reaction solution of the high blood lipid sample is turbid, add 550 μL of isopropanol to the reaction solution and mix thoroughly to dissolve the lipids, eliminate the turbidity, and measure the absorbance value. Calculate the absorbance by multiplying the measured absorbance by 1.5.
- 3、 Severe jaundice, hemolysis and other serum samples, it is necessary to prepare protein-free hemofiltrate (it is recommended that 100 μL of sample plus 300 μL of 5% trichloroacetic acid solution, mix thoroughly, equivalent to the sample diluted four times, 3000 rpm 4°C centrifugation for 10min, take the supernatant to be measured), and then calculate. Note the synchronous modification of the calculation formula.

Experimental example:

1. Take 100 μL of human serum sample, dilute 5 times with distilled water, follow the assay procedure, and measure with a 1 mL glass cuvette. Calculation: $\Delta A_t = A_t - A_b = 0.277 - 0 = 0.277$, $\Delta A_s = A_s - A_b = 0.472 - 0 = 0.472$ calculated by volume of liquid:

Glucose content (mg/mL) = $0.3125 \times \Delta A_t \div \Delta A_s \times 5 = 0.917$ mg/mL.

2. Take 0.1g of mouse muscle tissue, operate according to the steps of pre-treatment and assay, measured with 1mL glass cuvette Calculation: $\Delta A_t = A_t - A_b = 0.098 - 0 = 0.098$, $\Delta A_s = A_s - A_b = 0.472 - 0 = 0.472$, calculated by sample fresh weight:

Glucose content (mg/g) = $0.3125 \times \Delta A_t \div \Delta A_s \div W = 0.659$ mg/g.

References:

[1] Yee H Y, Goodwin J F. Evaluation of some factors influencing the o-toluidine reaction with glucose[J]. Analytical Chemistry, 2002, 45(13).DOI:10.1021/ac60335a030.

[2] Dubowski K M. An o-Toluidine Method for Body-Fluid Glucose Determination[J]. Clinical Chemistry, 1962(6) :6.DOI:10.1093/clinchem/8.6.592.

Related Products:

BC0340/BC0345	Glucogen Content Assay Kit
BC2540/BC2545	Cellulase(CL) Activity Assay Kit
BC0330/BC0335	Trehalose Content Assay Kit
BC2490/BC2495	Blood Glucose Content Assay Kit